



(An Autonomous Institution Under MHRD, Government of India)

Melakkottaiyur, Off Vandalur-Kelambakkam Road, Chennai – 600127, Tamil Nadu.

Ph : +91 44 2747 6317/6319

Fax : +91 44 2747 6301

Email : purchase@iitdm.ac.in

दिनांक /

Date: 22-4-24

PRE-BID MEETING

Sub: *Extension of Bid closing date and Pre Bid meeting minutes.*

3 No of Bidders participated in the Pre-Bid Meeting arranged on 05-04-2023 against our subject tender enquiry.

S.No.	Tender Specification	Amendment Request / Queries	Reply/Clarification
1.	<b>1. Body and Optics of microscope</b>  8. Motorized filter turret in 8 or more positions with individual filter cubes to image common Fluorophores (DAPI/Hoechst, GFP/FITC, RFP/TRITC, CFP).	Motorized filter turret in 6 or more positions with individual filter cubes to image common fluorophores (DAPI/Hoechst, GFP/FITC, RFP/TRITC, CFP).	<b>Request Accepted</b>  "Motorized filter turret in 6 or more positions with individual filter cubes to image common fluorophores (DAPI/Hoechst, GFP/FITC, RFP/TRITC, CFP)."  As many companies are having 6 position, very few companies only are having 8 position. We accept the request as it will not affect the performance.
2.	<b>2. Objective Lenses</b>  4. One Long working distance objective 40X, NA 0.6, WD 4.0 mm or longer	One Long working distance objective 40X, NA 0.6, WD 3.6 mm or longer	<b>We accept the request</b> , as it will not affect the performance.  "One Long working distance objective 40X, NA 0.6, WD 3.6 mm or longer"
3.	<b>6. Detection system for Epifluorescence Confocal and Widefield microscopy</b>  2. All the detectors should be directly connected to the scan head without any fibre optical interface to avoid signal loss and enhance the sensitivity of the system.	All the detectors should be directly/ separate connected to the scan head with/without any fibre optical interface to avoid signal loss and enhance the sensitivity of the system.	<b>We accept the request</b> , as fibre optical interface is shown to avoid losses and enhance the sensitivity of the system  "All the detectors should be directly/ separate connected to the scan head with/without any fibre optical interface to avoid signal loss and enhance the sensitivity of the system."
		The system should be capable	<b>We accept the request</b> as spectral

	<p><b>6. Detection system for Epifluorescence Confocal and Widefield microscopy</b></p> <p>4. The system should be capable of recording emission spectra with a spectral resolution of 2nm or better throughout the 400-800 nm spectral range. The spectral dispersion of the emission light should be based either on grating or prism with an ability to continuously adjust the emission bandwidth from either side of the spectrum down to 2nm.</p>	<p>of recording emission spectra with a spectral resolution of 5nm or better throughout the 400-800 nm spectral range. The spectral dispersion of the emission light should be based either on grating or prism with an ability to continuously adjust the emission bandwidth from either side of the spectrum down to 5nm.</p>	<p>resolution increase from 2 nm to 5 nm will not reduce the performance of the system for our applications.</p> <p>"The system should be capable of recording emission spectra with a spectral resolution of 5nm or better throughout the 400-800 nm spectral range. The spectral dispersion of the emission light should be based either on grating or prism with an ability to continuously adjust the emission bandwidth from either side of the spectrum down to 5nm."</p>
--	---	---	---

*[Handwritten signature]*

*ADD 2/24*

*[Handwritten signature]*  
Registrar

**Note:**  
 This Pre-Bid minutes shall form part of the tender document and the technical specification shall stand revised accordingly. The bidders are requested to note the change of technical specification and quote accordingly. **The Last date of submission of Bid is extended to 3.00 PM on 15.05.2024.**  
 The other terms and conditions of the tender document remains the same.

S/d  
Registrar

*ADD 2/24*